

屏東榮民總醫院病理檢驗科  
量測不確定度呈現一覽表

公告日期：2026.03.01

組別：生化免疫組(免疫定性)

評估期間：20250101-20251231

項目名稱	單位	目標平均值 (Target mean)	標準量測不確 定(SD)	相對標準量測不 確定(CV%)	擴充不確定度 (2*SD 或 2*CV%)	目標不確定度 (CVi 或 TEa)
Anti-HCV	S/CO	0.3				
	S/CO	5	0.27	5.41	10.82	Reactive
Anti-HBe	S/CO	2				
	S/CO	0.5	0.02	4.36	8.72	Reactive
Anti-HBs	mIU/mL	1				
	mIU/mL	15	1.14	7.52	15.04	Reactive
	mIU/mL	80	6.11	7.31	14.62	Reactive
Anti-HBc	S/CO	0.4				
	S/CO	2.73	0.12	5.22	10.44	Reactive
HBe Ag	S/CO	0.4				
	S/CO	4.05	0.4	10.93	21.86	Reactive
HBs Ag	S/CO	0.42				
	S/CO	3.5	0.28	8.27	16.54	Reactive
HIV Ag/Ab	S/CO	0.25				
	S/CO	6.35	0.56	10.63	21.26	Reactive
	S/CO	4.91	0.2	5.91	11.82	Reactive
	S/CO	3.23	0.19	6.76	13.52	Reactive
HAVAB IgG	S/CO	0.28				
	S/CO	2.28	0.11	4.86	9.72	Reactive
HAVAB IgM	S/CO	0.32				
	S/CO	1.88	0.13	7.49	14.98	Reactive
Anti-HBc IgM	S/CO	0.12				
	S/CO	3.22	0.22	7.43	14.86	Reactive
Rubella IgG	IU/mL	2				
	IU/mL	25	1.08	3.94	7.88	Reactive
	IU/mL	300	16.7	5.62	11.24	Reactive
SyphilisTP	S/CO	0.2				
	S/CO	2.5	0.35	13.96	27.92	Reactive

SD: Standard deviation, CV%: coefficient of variation, CVi: Individual biological variation, TEa: Total allowable error.

\*允收條件：擴充不確定度<目標不確定度。

填表人：



20260220

品質主管：



20260227 實驗室主管：



20260227